

Calibration information for ECO BB sensor

BBRT polynomial calculations for Seasave .con file							
See SBE app. Note No.87							
Sensor Wavelength = 660 nm							
User Inputs				27-Mar-14	17-Feb-16	NTU	
SN		053R	305R	305	053R		
Scale Factor	0.000003441	0.00323	0.00283619	0.003776	0.002963	Scale Factor	1.03
Dark Counts	50	0.142	0.0607	0.076	0.0439	Dark Counts	0.0607
	BBRT 305						
a0	-0.00017205	-0.00045866000	-0.00017215673	-0.00028697600	-0.00013007570		-0.06252100000
a1	0.000003441	0.00323000000	0.00283619000	0.00377600000	0.00296300000		1.03000000000
a2	0		0	0			0
a3	0		0	0			0