

Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 3213
CALIBRATION DATE: 12-Feb-12

SBE21 CONDUCTIVITY CALIBRATION DATA
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

GHIJ COEFFICIENTS

g = -4.21754306e+000
h = 4.97982351e-001
i = -3.31266329e-004
j = 4.34748749e-005
CPcor = -9.5700e-008 (nominal)
CTcor = 3.2500e-006 (nominal)

ABCDM COEFFICIENTS

a = 5.23250085e-006
b = 4.96865040e-001
c = -4.21206317e+000
d = -8.77310216e-005
m = 4.6
CPcor = -9.5700e-008 (nominal)

BATH TEMP (ITS-90)	BATH SAL (PSU)	BATH COND (Siemens/m)	INST FREQ (kHz)	INST COND (Siemens/m)	RESIDUAL (Siemens/m)
22.0000	0.0000	0.00000	2.91194	0.00000	0.00000
1.0000	34.9935	2.98966	8.27484	2.98968	0.00002
4.5000	34.9733	3.29807	8.63950	3.29804	-0.00003
15.0000	34.9308	4.28417	9.71270	4.28417	0.00000
18.5000	34.9208	4.63073	10.06219	4.63074	0.00001
23.9999	34.9091	5.19089	10.60214	5.19089	-0.00000
29.0001	34.9005	5.71456	11.08255	5.71456	-0.00000


Conductivity = $(g + hf^2 + if^3 + jf^4) / 10(1 + \delta t + \epsilon p)$ Siemens/meter

Conductivity = $(af^m + bf^2 + c + dt) / [10(1 + \epsilon p)]$ Siemens/meter

t = temperature[°C]; p = pressure[decibars]; δ = CTcor; ϵ = CPcor;

Residual = (instrument conductivity - bath conductivity) using g, h, i, j coefficients

Date, Slope Correction



● 11-Jun-10 0.9996777
▲ 12-Feb-12 1.0000000